

Application Data: Seeds (Moisture and Media Data)

Computrac® MAX® 1000/2000XL Moisture/Solids Analyzers

Material	Moisture	Test Time
Tomato, Hybrid Coated	0.8%	3–4 minutes
Tomato, Raw	8–9%	8–10 minutes
Celery	2–3%	2–3 minutes
Squash	4–5%	6–7 minutes
Cabbage	5–6%	6–7 minutes

Media Data

Material	Moisture	Test Time
Dry Paper	5–6%	2–3 minutes
Wet Paper	65–70%	5–6 minutes

Material	Moisture	Test Time
Sand	6–9 %	3–4 minutes
Soil	79-80%	15–20 minutes
Sand & Soil Mix	22–25%	7-8 minutes

The above sample results were achieved using rapid Loss-on-Drying methods. Some standard oven methods exceed 15 hours.



ARIZONA INSTRUMENT LLC

3375 N. Delaware Street, Chandler, AZ 85225 • (602) 470-1414 • (800) 528-7411 • www.azic.com

Exclusive Manufacturer of Computrac® Moisture Analyzers and Jerome® Toxic Gas Analyzers